



Unser Zeichen
25-P-Schu-0030

Bearbeiter
Kerstin Schulze
(Tel.: +493722 73 23-797)
(E-Mail: k.schulze@slg.eu)

Datum
03.07.2025

Notified Body Confirmation Letter

Reference: 25-P-Schu-0030

Confirmation of the status of a formal application, written agreement, and appropriate surveillance in the framework of Regulation EU 2023/607 amending Regulations (EU) 2017/745 and (EU) 2017/746 as regards the transitional provisions for certain medical devices and in vitro diagnostic medical devices

This letter confirms that, SLG Prüf- und Zertifizierungs GmbH, a Notified Body (NB) designated against Regulation (EU) 2017/745 (MDR) and identified by the number NB 0494 on NANDO, has received a formal application in accordance with Section 4.3, first subparagraph of Annex VII of MDR and has signed a written agreement in accordance with Section 4.3, second subparagraph of Annex VII of MDR with the following manufacturer:

Geratherm Respiratory GmbH

Kasernenstraße 4

97688 Bad Kissingen

SRN Number (if available): DE-MF-000006818



The devices covered by the formal application and the written agreement mentioned above are identified in the Tables below. Table 1 identifies the devices for which an MDR application has been received, written agreement concluded and for which the NB is also responsible for appropriate surveillance of the corresponding devices under the applicable Directive. Table 2 identifies the devices for which an MDR application has been received and a written agreement concluded, but the NB has not yet taken the responsibility for appropriate surveillance of the corresponding devices under the applicable Directive.

In the case of devices covered by certificates issued under Directive 90/385/EEC (AIMDD) or Directive 93/42/EEC (MDD) that expired after 26 May 2021 and before 20 March 2023, without having been withdrawn, this letter also confirms that the manufacturer signed the written agreement under MDR by the date of MDD/AIMDD certificate expiry; or provided evidence that a competent authority of a Member State had granted a derogation or exemption from the applicable conformity assessment procedure in accordance with Article 59(1) of MDR or Article 97(1) of the MDR respectively, by the 20 Mar 2023 for the relevant devices.

The transition timelines that apply to the devices covered by this letter, subject to the manufacturer's continued compliance to the other conditions specified in Article 120.3c of MDR (as amended by (EU) 2023/607), are shown below:

- 26 May 2026 for Class III custom-made implantable devices
- 31 December 2027 for Class III devices and Class IIb implantable devices excluding Well-established technologies (WET - sutures, staples, dental fillings, dental braces, tooth crowns, screws, wedges, plates, wires, pins, clips and connectors)
- 31 December 2028 for other Class IIb devices, Class IIa, Class I devices placed on the market in sterile condition or have a measuring function
- 31 December 2028 for devices not requiring the involvement of a notified body under MDD but requiring it under MDR (e.g., class I devices that qualify as re-usable surgical instruments)

This confirmation shall be renewed regularly and is valid until: 05.07.2026

On behalf of the Notified Body,



03.07.2025

Kerstin Schulze
Certification Body Medical Devices



Table 1: Devices covered by this letter and for which the NB is also responsible for appropriate surveillance of the corresponding devices under the applicable Directive:

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
Blue Cherry 406580340GR05AABYK	Class IIa	Blue Cherry	SLG/110168N2; NB0494
PFTstik, 4065803203GR03ADQW	Class IIa	Diffustik	SLG/110168N2; NB0494

**Confirmation Letter Revision History**

Date	NB internal reference traceable to each version of the letter	Action
2023-06-06	23-P-DEi-0023	Initial issue, 5003-23-PP-23-KP001a
2024-06-04	24-P-Schu-0036	Remove of MDR certified product, device 406580340GR05AABYK remaining (KP001b)
2025-07-03	25-P-Schu-0030	Regular renew, added of Diffustik upon request (it was part of 23-P-DEi-0023)